

VENTILATOR COMPATIBILITY LIST

KEY:	
•	Ventilator tested as compatible in this mode
0	Ventilator not completed validation in this mode

NOTE: $NOxBOX_i$ ® is not compatible with the use of combined inspiratory & expiratory limb ventilator circuits. The $NOxBOX_i$ ® device is available only pursuant to local laws and regulations.

 $NOxBOX_i$ ® is suitable for use on the currently approved patient population, which is neonates.

 $NOxBOX_i$ [®] is not intended to be used on other patient populations.

In some cases, a ventilator/gas delivery system is not able to pass criteria for the entire test procedure based on device design or specifications. NOxBOX, Ltd are able to supply formal reports detailing test conditions used upon request.

Manufacturer	Model	Neonatal	Pediatric	High Frequency	Transport	Anaesthesia	CPAP / High Flow Nasal Cannula
	Crossvent 2i+	•	•	0	•	0	O
Bio-Med Devices	MVP-10	•	•	0	•	0	0
Dio ivica Devices	TV-100	•	•	0	•	0	0
	LifePulse 203	•	0	1	0	0	0
Bunnell	LifePulse 204	•	0	•1	0	0	0
	Avea	•	0	0	0	0	0
	InfantFlow SiPAP	•	0	0	0	0	•
	LTV1000	0	•	0	0	0	0
CareFusion	LTV1200	0	•	0	0	0	0
	Revel	0	•	0	•	0	0
	Vela	0	•	0	0	0	0
CareFusion /	3100A	•	0	•	0	0	0
SensorMedics	3100B	0	•	•	0	0	0
	Apollo	•	•	0	0	•	0
	Babylog 8000 Plus	•	0	0	0	0	0
	Babylog VN500	•	0	0	0	0	0
Drägerwerk	Evita Infinity V500	•	•	0	0	0	0
· ·	Evita V600	•	•	0	0	0	0
	Evita V800	•	•	0	0	0	0
	Evita Infinity XL	0	•	0	0	0	0
Fisher & Paykel Healthcare	Optiflow Junior Breathing Circuit	0	0	0	0	0	•
Tieattiicare	Bubble CPAP	0	0	0	0	0	•
GE Healthcare	Carescape R860	•	•	0	0	0	0
	C1	•	•	0	0	0	0
	C2	•	0	0	0	0	0
	C3	•	•	0	0	0	0
Hamilton Medical	C6	•	•	0	0	0	0
	G5	•	•	0	0	0	0
	Galileo	0	•	0	0	0	0
	T1	•	•	0	•	0	0
IMT Medical (Vyaire)	Bellavista	•	•	0	0	0	0
	Servo-I	•	•	0	0	0	0
Maquet (Getinge)	Servo-N	•	•	0	0	0	0
maquet (Gettinge)	Servo-U	•	•	0	0	0	0
	Servo-Air	0	•	0	0	0	0
Novement (Covidion)	e360	0	•	0	0	0	0
Newport (Covidien)	HT70 Plus	0	•	0	•	0	0
Nihon Kohden	NKV-550	•	•	0	0	0	0

© NOxBOX Ltd

Unit 1 Eurolink Gateway, Castle Road, Sittingbourne, Kent, ME10 3AG, United Kingdom

Tel: +44 (0)1795 859430, Email: ask@noxboxltd.com

Registered Office: Forge, 43 Church Street West, Woking, Surrey, GU21 6HT, United Kingdom

Registered in: England and Wales. Registered No: 09563860

FRM-136.01 QA-12.04



VENTILATOR COMPATIBILITY LIST

KEY:	
•	Ventilator tested as compatible in this mode
0	Ventilator not completed validation in this mode

NOTE: *NOxBOX*_i® is not compatible with the use of combined inspiratory & expiratory limb ventilator circuits. The $NOxBOX_i^{\otimes}$ device is available only pursuant to local laws and regulations. $NOxBOX_i$ ® is suitable for use on the currently approved patient population, which is neonates. $NOxBOX_i$ [®] is not intended to be used on other patient populations. In some cases, a ventilator/gas delivery system is not able to pass criteria for the entire test procedure based on device design or specifications. NOxBOX, Ltd are able to supply formal reports detailing test conditions used upon request.

Manufacturer	Model	Neonatal	Pediatric	High Frequency	Transport	Anaesthesia	CPAP / High Flow Nasal Cannula
Percussionaire (Sentec)	VDR-4	•	•	•	0	0	0
Phillips Respironics	Esprit	0	•	0	0	0	0
Puritan Bennett	840	•	0	0	0	0	0
(Covidien)	980	•	0	0	0	0	0
Smiths Medical	VentiPac	0	•	0	•	0	0
Vapotherm	Precision Flow	0	0	0	0	0	•

© NOxBOX Ltd

Unit 1 Eurolink Gateway, Castle Road, Sittingbourne, Kent, ME10 3AG, United Kingdom

Tel: +44 (0)1795 859430, Email: ask@noxboxltd.com

Registered Office: Forge, 43 Church Street West, Woking, Surrey, GU21 6HT, United Kingdom

Registered in: England and Wales. Registered No: 09563860

FRM-136.01 QA-12.04

¹ Short term dose fluctuations can be observed; see set up guide for details